


Conformity and Technical Report -
#200420-01

Securing Lives

Because We Care!

Conformity and Technical Report - #200420-01 - BSM 5001

April 2020 - ©



This report includes the conformity and technical details of standard surgical gowns (Product ID BSM 5001) as of April 2020 produced in-house.

The company produces all of its goods in-house and currently exports to many countries including USA, Canada and EU.

OUR HEALTH MISSION

Our health mission is to supply the highest quality medical protective clothing while ensuring the highest customer satisfaction with continuous improvements and agile reliable product delivery.

We supply an extensive range of masks, aprons, scrubs, footwear and more, ideal for use in hospitals, clinics, GP surgeries, dental practices and other medical environments

BECAUSE WE CARE

The biggest threat in health care are pathogenic microorganisms – microorganisms can be transmitted with blood, body fluids, contaminated materials or medical instruments.

Medical protective clothing is essential equipment not only to protect medical professionals from pathogens, but also to protect patients from possible contamination by non-sterile garments.

STANDARD SURGICAL GOWNS



Produced in-house with CE - ISO 9001 - ISO 45001 and ISO 13485 certificates. The items can be shipped sterilized or non-sterilized depending on the requirements of the customer.

Material: 100% new pp non-woven or SMS non-woven

SMS Specifications: 3 layer structure, spunbond-meltblown-spunbond. Antistatic, breathable, low linting, Fluid repellent.

Weight: 36gsm - 43 gsm and 50 gsm

Applications: Surgical Gowns

Spunbond specifications: Hidrofobic, polipropilen non-woven fabric, one layer structure, 100% spunbond, Fluid repellent

Weight: 35 gsm

Applications: Patient and Visitor Gowns

Color: White - Blue - Green

Full length for maximum coverage. Latex free, sewn seams provide superior strength. Extra length on waist ties for easy securing in front. Additional ties at back of the neck and waist.


Type: Elastic wrist or with knitted cuff.

Sterilization method: EO Sterilization

Package Dimensions: 30x25x1,5 cm

Box Dimensions 60x40x40 / 100 pcs

Weight: Net 11 kg - Brut 11,50 kg

| STANDARD SURGICAL GOWN | Available Sizes | Dimensions |
|---|-----------------|--|
|  | S | A 112 cm B 103 cm C 29 cm D 54 cm E 135 cm |
| | M | A 114 cm B 105 cm C 29 cm D 56 cm E 140 cm |
| | L | A 116cm B 107cm C 29cm D 58cm E 145cm |
| | XL | A 118cm B 109cm C 29cm D 60cm E 150cm |
| | 2XL | A 120 cm B 111 cm C 29 cm D 62 cm E 155 cm |

STANDARD SURGICAL GOWNS

ACTUAL PRODUCT IMAGE



STANDARD SURGICAL GOWNS

OUR CERTIFICATES



SYNDICATE OF INTERNATIONAL SYSTEM CERTIFICATIONS



CE ATTESTATION OF CONFORMITY

Related Directives :
MEDICAL DEVICES 93/42/EEC-----TIBBİ CİHAZLAR DİREKTİFİ 93/42/EEC

Description of Product :
DISPOSABLE MEDICAL APRON
TEK KULLANIMLIK TIBBİ ÖNLÜK

Report Number /Rapor No :
M/TEC-4287 TEST VE KALİTE RAPORU

T rade M ar k

Regulations Applied acc. To Harmonized Standards:
EN14937, EN 61010-2-040, EN 61010-1, EN 60601-1-2 EMC, EN 60601-1

Manufactured by

Certificate No.: SISTURCE042020385
Issue Date (Original): 18.04.2020
Issue Date(Latest): 18.04.2021
Expiry Date: 17.04.2021



SYNDICATE OF INTERNATIONAL SYSTEM CERTIFICATIONS

This Certificate is issued under the following conditions:

- 1.It applies only to the above referenced models of the medical devices.
- 2.It does not imply that the SIS has performed any surveillance or control of their manufacture.
- 3.The manufacture is obligated to assure conformity of all in medical devices of the respective model to type assessed by the mean of this certificate.
- 4.The certificate remains valid until the manufacturing condition, the quality system or relevant legislation are changed .
- 5.After fulfilling of the relevant EU legislation requirements, the manufacture shall affix to each medical device, of the above referenced models, the CE-marketing according to this example:



Managing Director



Note: This certificate is valid only if produced with the continuation letter after the surveillance is carried out successfully.

The Organization's documentation and Implementation has been reviewed and found to comply with the relevant standard rules. This certificate of Registration is based on the evaluation of the mentioned scope given above. Organization is responsible for maintaining the responsibilities of the relevant standard rules. Any significant changes in the scope of the certification or standard referred above render this certificate invalid.

Corporate office(SIS):- Plot No. 1539, 2nd Floor, Sector-4,Gurgaon-122001, Haryana, India.
International office(SIS):- URB. Santa Ana Cal. German, Scherieber 276, San Isidro, Lima, Peru 15047.
Email us :-support@siscertifications.com, info@siscertifications.co.in. Call:- +91-9654721646
Web:- <http://www.siscertifications.co.in>, www.siscertifications.com
The status of this certificate can be verified on "<http://www.siscertifications.co.in>".

Issue No.: 01





Certificate

*Has been assessed and found to Comply with the Requirements of:
Denetlenmiş ve aşağıdaki standardın gerekliliklerine uygunluğu görülmüştür:*

ISO 45001:2018

*The Occupational Health and Safety Management System is applicable to:
İş Sağlığı Ve Güvenliği Yönetim Sistemi:*

**PRODUCTION AND SALES OF SHIRT, PROTECTIVE MASK, BONNET, OVERALLS,
BODY BAG, PROTECTIVE WORK DRESS, SURGICAL APRON**

**GÖMLEK, KORUYUCU MASKE, BONE, TULUM, CESET TORBASI,
KORUYUCU İŞ ELBİSESİ, CERRAHİ ÖNLÜK ÜRETİMİ VE SATIŞI**

Certificate No: ISO/21128
Sertifika Numarası: ISO/21128

Date of Issue: 18.04.2020
Yayınlanma Tarihi: 18.04.2020

1st Surveillance Audit : April 2021
1. Gözetim Denetim Tarihi : Nisan 2021
2nd Surveillance Audit : April 2022
2. Gözetim Denetim Tarihi : Nisan 2022

Date of Certification Audit: 10.04.2020
Denetim Tarihi: 10.04.2020

Date of Expiry: 17.04.2021
Son Geçerlilik Tarihi: 17.04.2021



QCS International Certifications Services PwL Ltd
Certificate of validity information E-mail: info@qcs-cert.com
mail to confirm



BIÇAKÇILAR LABORATUVAR MEDİKAL A.Ş.

Istiklal Mahallesi Atatürk Cad. No:21 Esenyurt İstanbul
Tel: +90 (212) 689 02 20 Fax: +90 (212) 689 02 29
http://www.bicakcilarlabmed.com.tr/
labinfo@bicakcilar.com

| | | |
|--|---|---------------------|
| Talep no: TLP – 12-666 Request no | HAVA GEÇİRGENLİK DENEY RAPORU TEST REPORT FOR AIR PERMEANCE REPORT | 00206.1303 rev01 |
| Teklif no: TLF – 12-175 Tender no | | 31.07.2012 |
| Kabul Tarihi: 11.06.2012 Receiving Date | | |
| Test başlangıç Tarihi: 12.06.2012 Test Initiation Date: | | |
| Test Bitiş Tarihi: 13.06.2012 Test Final Date: | | |

❖ MÜŞTERİ FİRMA- ADRES/ CUSTOMER-ADDRESS

❖ TEST AMACI / TEST TARGET

Numune ambalaj kağıtlarının hava geçirgenlik değerlerinin tespit edilmesi.
To determine the air permeance values for the packaging paper samples.

❖ NUMUNEALMA METODU / SAMPLING METHOD

Numuneler müşteri tarafından laboratuara gönderilmiştir.
Samples have been sent to the laboratory by the customer.
Numuneler standartta tanımlandığı şekilde, 10 cm x 10 cm boyutlarında kesilerek test edilmiştir.
Samples have been tested in accordance with the Standard method y cutting the samples as 10 cm x 10 cm dimensions.

❖ METOD TANIMI / METHOD DESCRIPTION

Hava Geçirgenlik Testi – BENDTSEN METODU
Air Permeance Test – BENDTSEN METHOD

❖ REFERANS STANDART/ REFERENCE STANDARD

ISO 5636-3:1992

❖ NUMUNE TANIMI – NUMUNE NO/ SAMPLE IDENTIFICATION – SAMPLE NO

Medikal Kraft Kağıt 50 x 100 mm
N12-0852

❖ NUMUNE SAYISI/ SAMPLE QUANTITY

10 Adet

❖ ÇEVRE ŞARTLARI/ TEST CONDITIONS

Sıcaklık/Temperature: 18-25 °C
Nem / Humidity: %30-60



❖ TAŞERON BİLGİLERİ/ SUBCONTRACTOR INFORMATION

Taşeron kullanılmamıştır/Not used

❖ ÖLÇÜM ANALİZ SONUÇLARI/ TEST RESULTS

| Numune no | Test Edilen Numune | Ölçüm Sonucu ml/dk | (µm/(Pa.s) cinsinden ölçüm sonucu | Aritmetik Ortalama | Standart Sapma |
|-----------|--------------------|--------------------|-----------------------------------|----------------------------|-------------------------------|
| 1 | KAĞIT | 900 | 10,17 | 914 ml/dk 10,32 µm/Pa.s | 53,995 ml/dk 0,611 µm/Pa.s |
| 2 | | 850 | 9,60 | | |
| 3 | | 900 | 10,17 | | |
| 4 | | 880 | 9,94 | | |
| 5 | | 850 | 9,60 | | |
| 6 | | 900 | 10,17 | | |
| 7 | | 980 | 11,07 | | |
| 8 | | 1000 | 11,30 | | |
| 9 | | 980 | 11,07 | | |
| 10 | | 900 | 10,17 | | |

❖ LİMITLER/ LİMİTS

EN 868-6: 2010 Packaging materials and systems for medical devices which are to be sterilized Part 6: Paper for the manufacture of packs for medical use for sterilization by ethylene oxide or irradiation- Requirements and test methods

Standardında tanımlı limit değerler: min. 0,2 □m/(Pa. s)

❖ ÖLÇÜM BELİRSİZLİĞİ/ MEASUREMENT UNCERTAINTY

98,4 ml/dk

❖ SAPMALAR/ DEVIATIONS

NA

❖ ONAYLAR/ APPROVALS

| | İsim/ Name | İmza/ Signature | Tarih/ Date |
|-------------------------------|---|-----------------|-------------|
| Testi Yapan/ Performed by: | Murat KAVAL KİMYA TEKN. | | 31.07.12 |
| Kontrol Eden / Controlled by: | Hüseyin AYDIN KİMYA MÜH. | | 31.07.12 |
| Lab. Müdürü/ Lab. Manager: | Ülkü ÖKER Yük.KİMYA MÜH. | | 31.07.12 |
| Mühür/Seal: | BIÇAKÇILAR LABORATUVAR MEDİKAL A.Ş. | | |

Test sonuçları bu test raporunda tanımlanan ve test edilen numuneler için geçerlidir.
Test results are valid only for the tested samples identified in this test report

Bu rapor Bıçakçılar Laboratuvarının yazılı onayı olmadan kısmen kopyalanamaz.
Imzasız ve mühürsüz raporlar geçersizdir.

This report shall not be reproduced other than in full except with the permission of Bıçakçılar Laboratory.
Testing reports without signature and seal are not valid

00206.1303 rev01
00206.1303
hazırlanmıştır. İbki rapor geçersizdir.

NA: Uygulanamaz/Not applicable



BIÇAKÇILAR LABORATUVAR MEDİKAL A.Ş.

İstiklal Mahallesi Atatürk Cad. No:21 Esenyurt İstanbul
Tel: +90 (212) 689 02 20 Fax: +90 (212) 689 02 29
http://www.bicakcilarlabmed.com.tr/
labinfo@bicakcilar.com

Talep No: TLP-12-1415
Request No

Teklif No:TLF-12-392
Tender No

Kabul Tarihi:22.11.2012
Receiving Date

Test Başlangıç Tarihi:23.11.2012
Test Initiation Date:

Test Bitiş Tarihi:07.12.2012
Test Final Date:



Test
TS EN ISO/IEC 17025
AB-0052-T

AB-0052-T

00206.2884

07.12.2012

STERİLİTE DENEY RAPORU (ISO 11737-2) STERILITY TEST REPORT

❖ MÜŞTERİ FİRMA- ADRES/ CUSTOMER-ADDRESS

❖ TEST AMACI / TEST TARGET

Bu testin amacı test edilen numunelerin sterilliğinin araştırılmasıdır.
The purpose of this test to investigate the test samples sterility.

❖ NUMUNEALMA METODU / SAMPLING METHOD

Müşteri tarafından gönderilmiştir.
It was sent by the customer.

❖ METOD TANIMI / METHOD DESCRIPTION

Ürün sterilitte testi LAF kabini altında direkt transfer metodu ile yapıldı.
Sterility test was performed under Laminar Air Flow by direct transfer method.

Test edilen ürün ya da SIP oranında kesilen numune, tüm numuneyi kapsayacak kadar TSB besiyeri ile dolu kavanozlara transfer edildi.
Sample or sample prepared as per SIP was transferred to the jar, filled with TSB growth medium enough to cover the sample.

Hazırlanan TSB kavanozları (30±2) °C'de inkübe edildi.
Prepared, TSB medium with product was incubated at (30±2)°C

Üreme yönünden, ürün içeren kavanozlar 14 gün boyunca kontrol edildi.
The jars containing the products were controlled for growth of microorganisms for 14 days.

❖ REFERANS STANDART/ REFERENCE STANDARD

ISO 11737-2:2009 Sterilization of Medical devices- Microbiological methods

❖ NUMUNE TANIMI – NUMUNE NO/ SAMPLE IDENTIFICATION – SAMPLE NO

Sudemed Markalı Sterilizasyon Rulosu İçindeki Spaç
N12-1854

❖ NUMUNE SAYISI/ SAMPLE QUANTITY

1 Adet

NA: Uygulanamaz/Not applicable

1 / 2

25.11.2011 / 260118-F07_03

**BIÇAKÇILAR LABORATUVAR MEDİKAL A.Ş.**

AB-0052-T

00206.2884

07.12.2012

❖ ÇEVRE ŞARTLARI/ TEST CONDITIONS

Sıcaklık/Temperature: 15-23 °C
Nem / Humidity: %35-60
Basınç Pressure: 0.02-0.08 inch H2O

❖ TAŞERON BİLGİLERİ/ SUBCONTRACTOR INFORMATION

Taşeron kullanılmamıştır / Not used

❖ NUMUNE ORANSAL PARÇASI/ SIP

NA

❖ TEST SONUÇLARI/ TEST RESULTS

| | TSB için for TSB |
|---|---------------------|
| Kontrol Edilen ürün sayısı Number of product controlled | 1 |
| Üreme görülen ürün sayısı ve gün Number and date of product growth observed | 0 |

❖ LİMİTLER/ LİMİTS

Müşteri tarafından belirlenir
Shall be determined by the customer

❖ ÖLÇÜM BELİRSİZLİĞİ/ MEASUREMENT UNCERTAINTY

Yoktur / Not present

❖ SAPMALAR/ DEVIATIONS

NA

❖ ONAYLAR/ APPROVALS

| | İsim/ Name | İmza/ Signature | Tarih/ Date |
|----------------------------|----------------------------|-----------------|-------------|
| Testi Yapan/ Performed by: | Havva CAVA BİYOLOG | | 07.12.12 |
| Uzman / Specialist: | Havva CAVA BİYOLOG | | 07.12.12 |
| Lab. Müdürü / Lab. Manager | Ülkü ÖKER Yük.KİMYA MÜH | | 07.12.12 |
| Mühür/Seal: | | | |

*Test sonuçları bu test raporunda tanımlanan ve test edilen numuneler için geçerlidir.
Test results are valid only for the tested samples identified in this test report
Bu rapor Biçakçılar Laboratuvarının yazılı onayı olmadan kısmen kopyalanamaz.
İmzasız ve mühürsüz raporlar geçersizdir.*

*This report shall not be reproduced other than in full except with the permission of Biçakçılar Laboratory.
Testing reports without signature and seal are not valid*

NA: Uygulanamaz/Not applicable

2 / 2

25.11.2011 / 260118-F07_03



SYNDICATE OF INTERNATIONAL SYSTEM CERTIFICATION

This Certificate has been awarded to

In recognition of the organization's Management System which complies with

ISO 9001:2015(QMS)

The scope of activities covered by this certificate is defined below

PRODUCTION AND SALES OF SHIRT, PROTECTIVE MASK, BONNET, OVERALLS, BODY BAG, PROTECTIVE WORK DRESS, SURGICAL APRON

GÖMLEK, KORUYUCU MASKE, BONE, TULUM, CESET TORBASI, KORUYUCU İŞ ELBİSESİ, CERRAHİ ÖNLÜK ÜRETİMİ VE SATIŞI

Certificate Number: **SISTURQ0420202037**
 Date of Issue of Original Certificate: **18.04.2020**
 Date of Issue of latest certificate: **18.04.2020**
 Expiry Date: **17.04.2021**

SYNDICATE OF INTERNATIONAL SYSTEM I

Managing Director



Note: This certificate is valid only if produced with the continuation letter after the surveillance is carried out successfully.

The Organization's documentation and Implementation has been reviewed and found to comply with the relevant standard rules. This certificate of Registration is based on the evaluation of the mentioned scope given above. Organization is responsible for maintaining the responsibilities of the relevant standard rules. Any significant changes in the scope of the certification or standard referred above render this certificate invalid. This is an accredited certificate issued by SIS Certifications Pvt. Ltd. sanctioned for issue by International Accreditation Services , 3060 Saturn Street Suite 100 Brea, California 92821-1732, USA.

Corporate office(SIS):- Plot No. 1539, 2nd Floor, Sector-4,Gurgaon-122001, Haryana, India.
 International office(SIS):- URB. Santa Ana Cal. German, Scherleber 276, San Isidro, Lima, Peru 15047.
 Email us :-support@siscertifications.com, info@siscertifications.co.in. Call:- +91-9654721646
 Web:- <http://www.siscertifications.co.in>, www.siscertifications.com
 The status of this certificate can be verified on "<http://www.siscertifications.co.in>".

Issue No.: 01

CERT.

S

EDITED
 ent Systems
 ation Body

MULTI
 AF
 ARABIAN

CERTIFICATE of Registration



*This is to Certify that the
Medical Devices – Quality Management System*

has been independently assessed and is compliant
with the requirements of

ISO 13485:2016

This Certificate is applicable to the following product or service ranges:

**PRODUCTION AND SALES OF SHIRT, PROTECTIVE MASK, BONNET,
OVERALLS, BODY BAG, PROTECTIVE WORK DRESS,
SURGICAL APRON**

**GÖMLEK, KORUYUCU MASKE, BONE, TULUM, CESET TORBASI,
KORUYUCU İŞ ELBİSESİ, CERRAHİ ÖNLÜK ÜRETİMİ VE SATIŞI**

:: Certificate No :: TR51935H

| | |
|--|---------------|
| Date of initial registration | 16 April 2020 |
| Date of this Certificate | 16 April 2020 |
| Surveillance audit on or before | 15 April 2021 |
| Recertification Due / Certificate expiry | 15 April 2023 |

This Certificate is property of Staunchly Management & System Services Ltd. and remains valid
subject to satisfactory surveillance audits.

Director

STAUNCHLY MANAGEMENT & SYSTEM SERVICES LTD.

Suite 48, 88-90 Hatton Garden, London, EC1N 8PN.

Phone : +44 345 680 0199

Email : info@staunchlyservices.com Web : www.staunchlyservices.com

For precise and updated information concerning the present certificate mail to info@staunchlyservices.com

This Certificate is the property of Staunchly Management & System Services Private Limited and shall be returned immediately when demanded



SMS/IF109A/17/REV02

**BIÇAKÇILAR LABORATUVAR MEDİKAL A.Ş.**

Kıraç Merkez Mahallesi Atatürk Cad. No.23 Büyükkçekmece 34909 İstanbul
Tel: +90 (212) 689 02 20 Fax: +90 (212) 689 02 29
<http://www.bicakcilarlabmed.com.tr/>
labinfo@bicakcilar.com

DENEY RAPORU
TESTING REPORT

00206.1302

14.06.2012

ÖLÇÜM ANALİZ SONUÇLARI / TEST RESULTS

| A-Kenarı | NUMUNE (N/15mm) | A-Kenarı | NUMUNE (N/15mm) |
|------------------|-----------------|----------|-----------------|
| 1 | 4,465 | 6 | 3,768 |
| 2 | 4,935 | 7 | 5,108 |
| 3 | 4,775 | 8 | 5,340 |
| 4 | 4,165 | 9 | 3,108 |
| 5 | 3,488 | 10 | 3,212 |
| ORTALAMA =4,2364 | | | |

| B-Kenarı | NUMUNE (N/15mm) | B-Kenarı | NUMUNE (N/15mm) |
|------------------|-----------------|----------|-----------------|
| 1 | 6,180 | 6 | 3,436 |
| 2 | 3,800 | 7 | 4,175 |
| 3 | 5,070 | 8 | 3,404 |
| 4 | 3,848 | 9 | 3,404 |
| 5 | 4,195 | 10 | 5,880 |
| ORTALAMA =4,3342 | | | |

LİMİT DEĞERLER/ LIMITS

EN868-5: 2009 a uygun olarak yapışma dayanımı min.1,5N/15mm olmalıdır

ÖLÇÜM BELİRSİZLİĞİ/ MEASUREMENT UNCERTAINTY

±0,04N

SAPMALAR/ DEVIATIONS

NA



BIÇAKÇILAR LABORATUVAR MEDİKAL A.Ş.

Kıraç Merkez Mahallesi Atatürk Cad. No:23 Büyükdere 34909 İstanbul
Tel: +90 (212) 689 02 20 Fax: +90 (212) 689 02 29
http://www.bicakcilarlabmed.com.tr
labinfo@bicakcilar.com

DENEY RAPORU TESTING REPORT

00206.1302

14.06.2012

| | Testi Yapan Performed by | Onaylayan Approved by | |
|--------------------------|---|---|---|
| Görev/Position | Uzman Teknisyen | Validasyon Kıdemli Uzmanı Validation Experienced Specialist | Lab. Müdürü Lab. Manager |
| Ad/Soyad Name/Surname | Çilem ORHAN | Özenç EFE ÖZTÜRK | Ülku ÖKER |
| İmza/Signature |  14.06.12 |  14.06.12 |  14.06.12 |
| Mühür/Seal |  | | |

Test sonuçları bu test raporunda tanımlanan ve test edilen numuneler için geçerlidir.
Bu rapor Bıçakçılar Laboratuvarının yazılı onayı olmadan kısmen kopyalanamaz. İmzasız ve mühürsüz raporlar geçersizdir.

Bu rapor Bıçakçılar Laboratuvarının yazılı onayı olmadan kısmen kopyalanamaz.
İmzasız ve mühürsüz raporlar geçersizdir.
This report shall not be reproduced other than in full except with the permission of Bıçakçılar Laboratory.
Testing reports without signature and seal are not valid

UYGUNLUK BEYANI

Belge No. M/TEC - 007
Rev. No. 0
Tarih: 01/ 01/ 2010

İşbu belge ile, aşağıda sıralanan ürünlerin

93/42/EEC Numaralı Tıbbi Cihazlar Yönetmeliğine

uygun olduğunu beyan ederiz.

Bu nedenle ürünler,

CE - uygunluk işaretinin

kullanılması ile ilgili tüm temel gerekli karşılamaktadır.

Sertifika No.: 022

Ürün Tipi: Sterilizasyon Ruloları ve Poşetleri
Buhar Biyolojik İndikatör
Sınıf: I
Marka Adı: AXIS

Karşılanan İlgili Uluslararası Standartlar

EN 867-5, EN 868-1&5, EN 11140-1, ISO 11607

İmalatçı:

İmza: —

Ürün Güvenliği :

ISO EN 13485

Geçerlilik Son Tarihi: 31.12.2014



BIÇAKÇILAR LABORATUVAR MEDİKAL A.Ş.

Kıraç Merkez Mahallesi Atatürk Cad. No:23 Büyükcemece 34909 İstanbul
Tel: +90 (212) 689 02 20 Fax: +90 (212) 689 02 29
<http://www.bicakcilarlabmed.com.tr/>
labinfo@bicakcilar.com

DENEY RAPORU TESTING REPORT

00206.1302

14.06.2012

| | | |
|---|---|--|
| Talep no: TLP – 12-666 Request no | Teklif no: TLF-12-175 Tender no | Raporun Sayfa Sayısı : 2 Total page number |
| Test Adı: Yapışma Sağlıklı Kuvveti Kontrolü Test Name: | Test Başlangıç Tarihi: 12.06.2012 Test Initiation Date | Test Bitiş Tarihi: 13.06.2012 Completion Date |

MÜŞTERİ BİLGİLERİ/SPONSOR INFORMATION

| | |
|------------------------------------|--|
| Müşteri Adı/ Customer Name | |
| Müşteri Adresi Customer Address | |

NUMUNE BİLGİLERİ/SAMPLE INFORMATION

| | |
|--|---|
| Numune Tanımı Sample Identification | N12-0852 Medikal Kraft Kağıt |
| Miktar Quantity | 10 Adet |
| Laboratuara ulaşma tarihi Sample receiving date | 11.06.2012 |
| Numune Alma Metodu Sampling Method | Numuneler müşteri tarafından laboratuara gönderilmiştir. Samples have been sent to the laboratory by the customer. |

TEST BİLGİLERİ/TEST INFORMATION

| | |
|---|---|
| Metod tanımı Method description | Herbir yapışma alanından, 15mm eninde ve cihaz aparatlarına uygun boyutta numune kesilir. Yapışma alanları numaralandırılırken ,herbir numune için altta kalan (CE işaretinin olduğu) kenar B olarak, karşısı ise A olarak tanımlanmıştır. Numune cihaz çeneleri arasına yerleştirilir ve 200mm/dk hızla çekilir. Maksimum kuvvet kaydedilir. |
| Referans standart no ve tarihi Standard number and date | EN868-5:2009 Annex D |
| Testin amacı Test target | Sıcaklıkla yapışabilir poşetlerin yapışma sağlıklı kuvvetinin belirlenmesi. |
| Malzemeler Materials | LF023 Gerilme Test Cihazı LF023-2 /5 Gerilme Test Cihazı Çeneleri |
| Ortam şartları Test conditions | Sıcaklık: 23±2 C° Nem : %40-%60 Rh |
| Şahit Numune Saklandı mı? Has witness sample been retained | Hayır |
| Varsa taşeron bilgileri Subcontractor information | NA |



BIÇAKÇILAR LABORATUVAR MEDİKAL A.Ş.

İstiklal Mahallesi Atatürk Cad. No:21 Esenyurt İstanbul
Tel: +90 (212) 559 02 20 Fax: +90 (212) 689 02 29
http://www.bicakcilarlabmed.com.tr
labinfo@bicakcilar.com

| |
|--|
| Talep no: TLP – 11-968 Request no: |
| Teklif no: TLF -11-237 Tender no: |
| Kabul Tarihi: 01.11.2011 Receiving Date: |
| Test başlangıç Tarihi: 03.11.2011 Test Initiation Date: |
| Test Bitiş Tarihi: 03.11.2011 Test Final Date: |



Test
TS EN ISO/IEC 17025
AB-0052-T

AB-0052-T

00036.12578

03.11.2011

PÜRÜZLÜLÜK DENEY RAPORU TEST REPORT FOR ROUGHNESS

❖ MÜŞTERİ FIRMA- ADRES/ CUSTOMER-ADDRESS

❖ TEST AMACI / TEST TARGET

Numune ambalaj kağıtlarının yüzey düzgünlük değerlerinin tespit edilmesi.
To determine the roughness values for the packaging paper samples.

❖ NUMUNEALMA METODU / SAMPLING METHOD

Numuneler müşteri tarafından laboratuvara gönderilmiştir.
Samples have been sent to the laboratory by the customer.

Numuneler standartta tanımlandığı şekilde, 75 mm x 75 mm boyutlarında kesilerek test edilmiştir.
Samples have been tested in accordance with the Standard method by cutting the samples as 75 mm x 75 mm dimensions.

❖ METOD TANIMI / METHOD DESCRIPTION

Pürüzlülük Testi – BENDTSEN METODU
Roughness Test – BENDTSEN METHOD

❖ REFERANS STANDART/ REFERENCE STANDARD

ISO 8791-2

❖ NUMUNE TANIMI – NUMUNE NO/ SAMPLE IDENTIFICATION – SAMPLE NO

Sterilizasyon Rulosu
N11-1795

❖ NUMUNE SAYISI/ SAMPLE QUANTITY

1 Rulo

❖ ÇEVRE ŞARTLARI/ TEST CONDITIONS

Sıcaklık/Temperature: 21-25 °C
Nem / Humidity: %30-60

NA: Uygulanamaz/Not applicable

1 / 2

01.03.11 / 270118-F42-00

**BIÇAKÇILAR LABORATUVAR MEDİKAL A.Ş.**

AB-0052-T

00036.12578

03.11.2011

❖ TAŞERON BİLGİLERİ/ SUBCONTRACTOR INFORMATION

Taşeron kullanılmamıştır/Not used

❖ ÖLÇÜM ANALİZ SONUÇLARI/ TEST RESULTS

| Numune no | Test Edilen Numune | Ölçüm Sonucu ml/dk | Aritmetik Ortalama | Standart Sapma |
|-----------|--------------------|--------------------|--------------------|----------------|
| 1 | KAGIT | 380 | 376,5 ml/dk | 14,53922 ml/dk |
| 2 | | 365 | | |
| 3 | | 370 | | |
| 4 | | 370 | | |
| 5 | | 380 | | |
| 6 | | 400 | | |
| 7 | | 360 | | |
| 8 | | 400 | | |
| 9 | | 360 | | |
| 10 | | 380 | | |

❖ LİMITLER/ LIMITS

EN 868-6: 1997 Packaging materials and systems for medical devices which are to be sterilized Part 6: Paper for the manufacture of packs for medical use for sterilization by ethylene oxide or irradiation-Requirements and test methods

Standartında tanımlı limit değerler: min.(50ml/dk)

max.(1200 ml/dk)

❖ ÖLÇÜM BELİRSİZLİĞİ/ MEASUREMENT UNCERTAINTY

45,5 ml/dk

❖ SAPMALAR/ DEVIATIONS

NA

❖ ONAYLAR/ APPROVALS

| | | | |
|-------------------------------------|-------------|-----------------|-------------|
| Testi Yapan/ Performed by: | İsim/ Name | İmza/ Signature | Tarih/ Date |
| | Murat KAVAL | | 03.11.2011 |
| Teknik Yönetici/ Technical Manager: | Melda AYDIN | | 03.11.11 |
| Lab. Müdürü / Lab. Manager | Ölkü ÖKER | | 03.11.11 |
| Mühür/Seal: | | | |

Test sonuçları bu test raporunda tanımlanan ve test edilen numuneler için geçerlidir.
Test results are valid only for the tested samples identified in this test report

Bu rapor Biçakçılar Laboratuvarının yazılı onayı olmadan kısmen kopyalanamaz.
İmzasız ve mühürsüz raporlar geçersizdir.
This report shall not be reproduced other than in full except with the permission of Biçakçılar Laboratory.
Testing reports without signature and seal are not valid

| | | | |
|--|--|----------|------------|
| | | DATE | 19.03.2014 |
| | | DOC.NO | F32 |
| | | PAGE NO | 1 |
| | | REV.NO | 0 |
| | | REV.DATE | 00.01.1900 |

TECHNICAL DATA SHEET

Product : Polypropylene

Product Description : ENDLESS FILAMENTS MELTBLOWN, THERMALLY BONDED.
Raw Material : 100 % PP
Application on Fabric : SMS /HYDROPHOBIC
Treatment :
Fabric Colour : MEDICAL BLUE
Customer Name :
Weight : 40 GSM
Width :
Packing : PE BAG WITH LABEL

| PROPERTIES | TEST METHOD | UNIT | TARGET |
|---------------------|--------------------|--------|--------|
| WEIGHT | NWSP 130.1.R0 (15) | gsm | 40 |
| THICKNESS | NWSP 120.1.R0 (15) | mm | 0.36 |
| TENSILE STRENGTH | MD | N/S cm | 108.0 |
| | CD | | 50.0 |
| ELONGATION AT BREAK | MD | % | 136.0 |
| | CD | | 119.5 |
| HYDROSTATIC HEAD | NWSP 080.6.R0(15) | mm | 400.0 |

Tolerances For The Average Results

| | | |
|----------------------------|---------|---|
| Weight | ± 5 % | Roll Tolerance Length : - 0 / +5% against target / ordered length Width : Up to 150 cm in width = -0mm/+5mm Over 150 cm in width = - 0mm/+10 Splice : Maxium five splices per roll |
| Thickness | ± 10 % | |
| Tensile Strenght | ± 15 % | |
| Elongation | ± 15 % | |
| Hydrostatic Head | ± 15 % | |
| Liquid Strike-Through Time | ± 0,5 % | |
| Air Permeability | ± 20 % | |
| Absorption | ± 20 % | |

The product is wound onto cardboard cores and then wrapped in polyethylen film. Bar code labels with product code, description and lot details are applied to the outside of each pack and a small label is applied to each roll. Suitable sized rolls may be palletised and the pallet load cling-wrapped

| | |
|------------------|--------------------------|
| Preparation Date | QUALITY CONTROL APPROVAL |
| 06.04.2020 | |



**EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.**
Esenyurt Firuzköy Bulvarı No:29 34325 Avcılar
İstanbul/ TÜRKİYE

TEST REPORT
DENEY RAPORU

20009637

03-20

EKOTEKS

Customer name:

Address:

Buyer name:

Contact Person:

Order No:

-

Article No:

SMS FABRIC

Name and identity of test item:

White non-woven fabric. (Claimed to be; 100 % PP , 30 gsm , Color Code: White)

The date of receipt of test item:

09.03.2020

Re-submitted/re-confirmation date:

-

Date of test:

09.03.2020-13.03.2020

Remarks:

-

Sampling:

The results given in this report belong to the received sample by vendor.

End-Use:

-

Care Label:

Not Specified

Number of pages of the report: 4

Seal

Date
13.03.2020

Customer Representative
Hatice ACARALP

Head of Testing Laboratory
Sevim A. RAZAK
13.03.2020

This report shall not be reproduced other than in full except with the permission of the laboratory.
Testing reports without signature and seal are not valid.

**EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.**

20009637

03-20

| REQUIRED TESTS | RESULT | COMMENTS |
|--|--------|----------|
| MICROBIOLOGICAL TEST | | |
| Antibacterial Activity | P | |
| P: Pass F: Fail R: Refer to retailer technologist. | | |

Test results were evaluated according to EKOTEKS 70 (REF: 14683:2014) requirements

REMARK: Original samples are kept for 3 months and all technical records are kept for 5 years unless otherwise specified. If requested, measurement uncertainty will be reported. But unless otherwise specified, measurement uncertainty is not considered while stating compliance with specification or limit values. The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a level of confidence of approximately 95%. Tests marked (*) in this report are not included in the accreditation schedule.

Gen. 1136-/03



This report shall not be reproduced other than in full except with the permission of the laboratory.
Testing reports without signature and seal are not valid.

**EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.**

20009637

03-20

TEST RESULT

BACTERIAL FILTRATION EFFICIENCY (BFE)

Test Metod: EKOTEKS 70 (In-House Method-Bacterial Filtration Efficiency Testing /Ref: TS EN 14683:2019 Medical Face Masks, Requirements and Test Methods (*))

A specimen of the mask material is clamped between a impactor and an aerosol chamber. An aerosol of *Staphylococcus aureus* is introduced into the aerosol chamber and drawn through the mask material and the impactor under vacuum. The bacterial filtration efficiency of the mask is given by the number of colony forming units passing through the medical face mask material expressed as a percentage of the number of colony forming units present in the challenge aerosol.

| | |
|---|--|
| Test Flow Rate | 28,3 L/min |
| Test Flow Time | 2 minute |
| Sample Sizes | 5 piece (10x10 cm ²) |
| Microorganism | <i>Staphylococcus aureus</i> ATCC 6538 |
| Bacterial concentration (cfu/ ml) | 5x10 ⁵ cfu/ ml |
| incubation conditions | 24 hour, 35°C ± 2°C |
| Positive control average(C) | 3.0x10 ³ cfu/ ml |

RESULTS

| Test Sample (T) | Number of Bacteria (cfu/ml) | Bacterial Filtration Efficiency (% B) |
|-----------------|-----------------------------|---|
| 1 | 115 | 96.2 % |
| 2 | 110 | 96.3 % |
| 3 | 120 | 96.0 % |
| 4 | 126 | 95.8 % |
| 5 | 130 | 95.7 % |

cfu: Colony-forming unit

$$B = (C - T) / C \times 100$$

%B: Bacterial Filtration Efficiency

C: is the mean of the total plate counts for the two positive control runs

T: is the total plate count for the test specimen